SECTION 1 IDENTIFICATION

- (a) Product Identifier: Clear Image® Ultrasound Scanning Gel
- (b) Synonyms: ultrasound couplant, scanning gel, ultrasound gel, coupling media
- (c) Use: Ultrasound coupling media for diagnostic imaging; non-sterile; for external use only. (Not for use with defibrillators.)
- (d) Manufacturer: NEXT Medical Products Company

45 Columbia Road, Branchburg, NJ 08876

800-458-4254

(e) Emergency Phone Number: 800-458-4254 (8:30am-5pm Eastern Time)

SECTION 2 HAZARD(S) IDENTIFICATION

(a) Not classified as a hazardous material

(b) Signal word: Not applicable(c) Eyes: Transient irritationSkin: No data availableInhalation: None expected

Ingestion: None expected

Note: Allergic reactions are rare, but may occur with individuals sensitive to one or more

ingredients.
(d) Not applicable

SECTION 3 COMPOSITION/INFORMATION ON INGREDIENTS

Materials not considered to be hazardous under OSHA's Hazard Communication Standard (29 CFR 1900.1200). Product does not contain any ingredients having known health hazards in concentrations \geq 1%, or any known or suspected carcinogens in quantities \geq 0.1%.

SECTION 4 FIRST-AID MEASURES

(a) Eyes: Flush with water for at least 15 minutes

Skin: Remove with damp towel Inhalation: not applicable

Ingestion: Seek immediate medical attention if large quantity is ingested and discomfort occurs

- (b) Important symptoms: not applicable
- (c) Indication of immediate medical attention: No further relevant information available

SECTION 5 FIRE-FIGHTING MEASURES

- (a) Extinguishing media: all types
- (b) Specific hazards: none
- (c) Special equipment precautions: none

SECTION 6 ACCIDENTAL RELEASE MEASURES

- (a) Precaution: Material is slippery.
- (b) Method for cleaning up: Scoop or wipe up excess. Wash area with water. Sprinkle area with traction material if spill not cleaned immediately.

SECTION 7 HANDLING AND STORAGE

- (a) Precautions: not applicable
- (b) Conditions: Store near room temperature away from sunlight to preserve viscosity.

SECTION 8 EXPOSURE CONTROLS/PERSONAL PROTECTION

- (a) PEL limit: not applicable
- (b) Engineering controls: not applicable
- (c) Individual protection measures: protective gloves on individuals demonstrating sensitivity to this product.

SECTION 9 PHYSICAL AND CHEMICAL PROPERTIES

- (a) Appearance: Water white, opaque gel
- (b) Odor: bland
- (c) Odor threshold: not applicable
- (d) pH: 6.0 8.0
- (e) Melting point/freezing point: No data available
- (f) Initial boiling point and boiling range: No data available
- (g) Flash point: No data available
- (h) Evaporation rate: No data available
- (i) Flammability: No data available
- (j) Upper/lower flammability or explosive limits: not applicable
- (k) Vapor Pressure: not determined
- (I) Vapor density: not determined
- (m) Relative density: not determined
- (n) Solubility(ies): Complete solubility in water
- (o) Partition coefficient: n-octanol/water: not determined
- (p) Auto-ignition temperature: not determined
- (g) Decomposition temperature: not determined
- (r) Viscosity: Available in standard high viscosity and medium viscosity versions

SECTION 10 STABILITY AND REACTIVITY

- (a) Reactivity: Not applicable
- (b) Chemical Stability: Stable
- (c) Possibility of hazardous reactions: None known
- (d) Conditions to avoid: None known
- (e) Incompatibility: None known
- (f) Hazardous Decomposition Products: None known

SECTION 11 TOXICOLOGICAL INFORMATION

- (a) Route of exposure: No data available
- (b) Symptoms: None known
- (c) Delayed and immediate effects: None known
- (d) Numerical measures: None known
- (e) Listing in NTP or IARC or OSHA: Not applicable

SECTION 12 ECOLOGICAL INFORMATION

No data available

SECTION 13 DISPOSAL CONSIDERATIONS

Follow applicable federal, state and local regulations

SECTION 14 TRANSPORT INFORMATION

Not applicable

SECTION 15 REGULATORY INFORMATION	
-----------------------------------	--

All ingredients are listed on the TSCA inventory.

SECTION 16 OTHER INFORMATION

Date of preparation: November 22, 2022.

Revision Level: E (formerly 70054D 5/4/2015, 70054C 12/27/13; 70054B 1/8/14; 70054A 10/16/12; Magnaflux 7/5/11)

This SDS has been prepared in compliance with the Federal OSHA Hazard Communication Standard, 29 CFR 1910.1200 and HCS June 1, 2015. The information contained herein is accurate to the best of our knowledge as of the date of this publication. NEXT Medical Products makes no warranty of any kind, express or implied, concerning the safe use of this product.